Carestream

DECLARATION OF CONFORMITY

Carestream Health, Inc., hereby declares under its sole responsibility that the product(s) listed are made in conformity with ANNEX I, Essential Requirements and ANNEX VII, EC Declaration of Conformity, of the European Economic Community Medical Device Directive [Directive 93/42/EEC]; Article 4 of the Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment [Directive 2011/65/EU]; ANNEX II of the Radio Equipment Directive [Directive 2014/53/EU]; and the requirements of Clause 6.6 of Schedule 3 of the Australian Therapeutic Goods (Medical Devices) Regulations 2002 relating to the stated devices.

Manufacturer's Name and Address:	Carestream Health, Inc. 150 Verona Street Rochester, New York USA 14608
Medical Device:	Dental Video Camera
Product List:	CS 1500, Wireless version "End of List"
Device Classification:	Class I, Rule 5 (Council Directive 93/42/ EEC, ANNEX IX) Class I, Schedule 2, Part 4, Rule 4.1 (Australian Therapeutic Goods (Medical Devices) Regulations 2002)
GMDN Code and Term:	45099, Flexible video stomatoscope
Scope of Application:	All Declared Products

Each kind of medical device to which the Declaration of Conformity (not requiring assessment by the Secretary) procedures have been applied complies with the applicable provisions of the essential principles and the classification rules before being supplied.

European Authorized Representative:

Carestream Health France 1, rue Galilée 93192 NOISY-LE-GRAND CEDEX FRANCE

Standards Applied

EN ISO 13485: 2012	Medical devices – Quality management systems – Requirements for regulatory purposes
EN ISO 14971:2012	Medical devices - Application of risk management to medical devices
EN 980:2008	Symbols for use in the labeling of medical devices
EN 1041:2008	Information supplied by the manufacturer of medical devices
EN 60601-1: 2006/A1: 2013	Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
EN 60601-1-2: 2007/AC: 2010	Medical electrical equipment Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
EN 60601-1-6: 2010	Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability
EN 62366: 2008	Medical devices - Application of usability engineering to medical devices
EN 62304: 2006/AC: 2008	Medical device software – Software life-cycle processes
EN 62471: 2008	Photobiological safety of lamps and lamp systems
EN ISO 10993-1: 2009/AC: 2010	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
EN 300 328 V2.1.1:2016	Electromagnetic compatibility and Radio Spectrum Matters (ERM): Wideband Transmission systems; Data transmission equipment operating in the 2.4 GHz ISM band and using wide band modulation techniques; Harmonized Standard covering the essential requirements of article 3.2 of the Directive 2014/53/EU
EN 62479:2010	Assessment of the compliance of low power electronic and electrical equipment with the basic restrictions related to human exposure to electromagnetic fields (10 MHz to 300 GHz)
EN 62133: 2013	Secondary cells and batteries containing alkaline or other non-acid electrolytes. Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications
EN 50581: 2012	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

EN 62321: 2009

EN 62474: 2012

Electrotechnical products – Determination of levels of six regulated substances (lead, mercury, cadmium, hexavalent chromium, polybrominated biphenyls, polybrominated diphenyl ethers)

Material declaration for products of and for the electrotechnical industry

Kevin Wright Senior Director World Wide Regulatory Affairs Carestream Health, Inc. 150 Verona Street Rochester, New York USA 14608 Telephone 585-627-6528